



HAL
open science

Reproducibility in Echocardiographic Assessment of Diastolic Function in a Population Based Study (The STANISLAS Cohort Study)

Zied Frikha, Nicolas Girerd, Olivier Huttin, Pierre-Yves Courand, Erwan Bozec, Arnaud Olivier, Zohra Lamiral, Faiez Zannad, Patrick Rossignol

► **To cite this version:**

Zied Frikha, Nicolas Girerd, Olivier Huttin, Pierre-Yves Courand, Erwan Bozec, et al.. Reproducibility in Echocardiographic Assessment of Diastolic Function in a Population Based Study (The STANISLAS Cohort Study). PLoS ONE, 2015, 10 (4), 10.1371/journal.pone.0122336 . hal-01758213

HAL Id: hal-01758213

<https://hal.univ-lorraine.fr/hal-01758213>

Submitted on 4 Apr 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

RESEARCH ARTICLE

Reproducibility in Echocardiographic Assessment of Diastolic Function in a Population Based Study (The STANISLAS Cohort Study)

Zied Frikha¹, Nicolas Girerd¹, Olivier Huttin², Pierre Yves Courand^{3,4,5}, Erwan Bozec¹, Arnaud Olivier², Zohra Lamiral¹, Faiez Zannad¹, Patrick Rossignol^{1*}

1 INSERM, Centre d'Investigations Cliniques 9501, Université de Lorraine, CHU de Nancy, Institut Lorrain du cœur et des vaisseaux, Nancy, France, **2** Service de Cardiologie, Institut Lorrain du Cœur et des Vaisseaux, Centre Hospitalier Universitaire de Nancy, Nancy, France, **3** Fédération de cardiologie, European Society of Hypertension Excellence Centre, hôpital de la Croix-Rousse, hospices civils de Lyon, 103, Grande Rue de la Croix-Rousse, 69004 Lyon, France, **4** Génomique fonctionnelle de l'hypertension artérielle, EA 4173, université Claude-Bernard Lyon 1, 69100 Villeurbanne, France, **5** Hôpital Nord-Ouest, 69400 Villefranche-sur-Saône, France

* p.rossignol@chu-nancy.fr



OPEN ACCESS

Citation: Frikha Z, Girerd N, Huttin O, Courand PY, Bozec E, Olivier A, et al. (2015) Reproducibility in Echocardiographic Assessment of Diastolic Function in a Population Based Study (The STANISLAS Cohort Study). PLoS ONE 10(4): e0122336. doi:10.1371/journal.pone.0122336

Academic Editor: Wen-Chih Hank Wu, Providence VA Medical Center and Brown University, UNITED STATES

Received: October 28, 2014

Accepted: February 10, 2015

Published: April 8, 2015

Copyright: © 2015 Frikha et al. This is an open access article distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

Data Availability Statement: Due to legal restrictions, data are available upon request. Interested researchers may request access to data upon approval from the Stanislas scientific board of the study. This scientific board can be reached through the primary investigator of the study, Pr Faiez Zannad (f.zannad@chu-nancy.fr).

Funding: This study was sponsored by Nancy CHU and granted by the Lorraine Region, the Programme Hospitalier de Recherche Clinique Interregional (BioSe), the 7th EU-FP MEDIA project (FP7)

Abstract

Introduction

There is limited evidence regarding intra-observer and inter-observer variations in echocardiographic measurements of diastolic function. This study aimed to assess this reproducibility within a population-based cohort study.

Methods

Sixty subjects in sinus rhythm were randomly selected among 4th visit participants of the STANISLAS Cohort (Lorraine region, France). This 4th examination systematically included M-mode, 2-dimensional, DTI and pulsed-wave Doppler echocardiograms. Reproducibility of variables was studied by intra-class correlation coefficients (ICC) and Bland Altman plots.

Results

Our population was on average middle-aged (50 ± 14 y), overweight ($BMI = 26 \pm 6 \text{ kg/m}^2$) and non-smoking (87%) with a quarter of the participants having self-declared hypertension or treated with anti-hypertensive medication(s). Intra-observer ICC were > 0.90 for all analyzed parameters except for left ventricular ejection fraction (LVEF) which was 0.89 (0.81–0.93). The mean relative intra-observer differences were small and limits of agreement of relative differences were narrow for all considered parameters ($< 5\%$ and $< 15\%$ respectively). Inter-observer ICC were > 0.90 for all analyzed parameters except for LVEF (ICC = 0.87) and both mitral and pulmonary A wave duration (0.83 and 0.73 respectively). The mean relative inter-observer

#261409), the HOMAGE (FP7 #305507) and the FIBRO-TARGETS project (FP7 #602904). The funders had no role in study design, data collection and analysis, decision to publish, or preparation of the manuscript.

Competing Interests: The authors have declared that no competing interests exist.

differences were <5% for all parameters except for pulmonary A wave duration (mean difference = 6.5%). Limits of agreement of relative differences were narrow (<15%), except for mitral A wave duration and velocity (both <20%) as well as left ventricular mass and pulmonary A wave duration (both <30%). Intra-observer agreements with regard to the presence and severity of diastolic dysfunction were excellent (Kappa = 0.93 (0.83–1.00) and 0.88 (0.75–0.99), respectively).

Conclusion

In this validation study within the STANISLAS cohort, diastolic function echocardiographic parameters were found to be highly reproducible. Diastolic dysfunction consequently appears as a highly effective clinical and research tool.

Introduction

Heart failure (HF) with preserved left ventricular (LV) ejection fraction (EF) is a progressive disorder characterized by impaired LV relaxation, increased LV stiffness, increased interstitial deposition of collagen and modified extracellular matrix proteins. HF with preserved LVEF (HFPEF) is frequently referred to as diastolic heart failure because of the presence of diastolic LV dysfunction resulting from slow LV relaxation and increased LV stiffness [1]. Because the process of LV myocardial remodeling invariably begins before the onset of symptoms, current guidelines place special emphasis on the detection of subclinical diastolic dysfunction and the timely identification of patients who are at risk of developing overt HF [2].

The gold standard for assessing diastolic function remains pressure-volume relationship. However, this invasive approach is never used in clinical practice. An extensive non-invasive assessment of diastolic function is now possible, and its clinical evaluation is mostly based on echocardiographic parameters [3]. Indeed, echocardiography provides valuable information regarding LV diastolic function in addition to classical M-mode and 2D echocardiography and pulsed-wave Doppler. Nevertheless, these techniques remain complex since no single measurement reflects diastolic function. Thus, a comprehensive assessment of a number of variables is required to evaluate diastolic function as accurately as possible [3–5].

The value of a given parameter is based on its ability and reproducibility to correctly predict a clinical entity. Therefore, each parameter of diastolic function should be obtained in a standardized and reliable manner.

Verifying the reliability of each parameter of diastolic function is a crucial aspect in clinical practice because it assists clinicians in establishing the diagnosis of HFPEF and decision-making especially in situations of discordant echocardiographic parameters. In epidemiology, accuracy of a measurement process is defined as the degree to which a measurement represents the true value of the parameter being measured. Reproducibility refers to the capacity to produce the same result on each occasion under identical conditions [6]. A measurement process is accurate, or unbiased, if the expected value of the measurement is the true value of the parameter being estimated. Presently, available results from other studies on intra-observer and inter-observer variations of all echocardiographic measurements for diastolic function are limited.

Recently, an array of HF biomarkers has emerged, each reflecting different pathophysiological processes in the development and progression of HF, namely myocardial insult, inflammation and remodeling. The assessment of a new biomarker should be made on a wide range of patients and in typical conditions of its intended use. The diagnostic methods used to detect

diastolic HF should thus be contemporary, rigorous, standardized and fair. Consequently, we endeavored to assess LV diastolic function using an accurate measurement process.

The present study was aimed to validate a standardized process of echocardiographic assessment of diastolic function in a population-based study by analyzing both intra-observer and inter-observer reproducibility of all parameters of diastolic function.

Methods

Study population

The STANISLAS Cohort is a single-center familial longitudinal cohort comprised of 1006 families (4295 subjects) from the Nancy region recruited in 1993–1995 at the Centre for Preventive Medicine. This cohort was established with the primary objective of investigating gene-gene and gene-environment interactions in the field of cardiovascular diseases. The families were deemed healthy, free of declared acute and/or chronic illness, in order to assess the effect of genetics on the variability of intermediate phenotypes on the transition toward pathology.

From 2011 to 2014, 1200 survivors of the original cohort underwent their 4th examination at our department. M-mode, 2-dimensional, DTI and pulsed-wave Doppler echocardiograms were routinely performed at the 4th examination. For the present study, 60 subjects in sinus rhythm were randomly selected from the STANISLAS Cohort study.

The research protocol was approved by local Ethics Committee (Comité de Protection des Personnes Est III—Nancy—France) and all study participants gave a written informed consent to participate. The informed written consent was previously approved by the local Ethics Committee.

Study design

Echocardiographic image acquisition. Examinations were performed by an experienced echocardiographer in the left lateral decubitus position with a commercially available standard ultrasound scanner (Vivid 9, General Electric Medical Systems, Horten, Norway) using a 2.5 MHz phased-array transducer (M5S). The echo/Doppler examination included parasternal long- and short axis views and three standard apical views. For each view, at least three consecutive cardiac cycles were recorded during quiet respiration.

From the three apical views, separate M-mode and color tissue Doppler acquisitions were recorded. The Doppler pulse repetition frequency was 1 kHz. Pulsed-wave (PW) Doppler was performed in the apical 4-chamber view to obtain mitral inflow velocities to assess LV filling. Recording for isovolumic relaxation time measurement was obtained by simultaneous recording of the aortic and mitral flows. PW Doppler Tissue Imaging (DTI) was performed in the apical views to acquire mitral annular velocities. The sample volume was positioned at 1 cm of the septal and lateral insertion sites of the mitral leaflets and adjusted as necessary (usually 5–10 mm) to cover the longitudinal excursion of the mitral annulus in both systole and diastole. Pulmonary venous flow velocities were recorded into the right upper pulmonary vein from the apical 4-chamber view. Color flow imaging was used for the proper location of the sample volume and the 2-mm sample volume was placed 5 mm into the pulmonary vein for optimal recording of the spectral waveforms.

All acquired images and media were stored on a secured network server as digital clips, using a unique identification number and analyzed on a dedicated workstation (EchoPAC PC, version 110.1.0, GE Healthcare).

Reproducibility analysis. Two experienced physician echocardiographers (Z.F. and O.H.), blinded to each other and to the demographic parameters, performed a separate complete analysis of diastolic function of the 60 participants. The first physician echocardiographer (Z.F.) re-analyzed

the entire data set of all 60 study patients a second time. The re-analyses were performed in random order after a period of 4 weeks. The first reading of the first analyzer (Z.F.) was considered for the statistical analysis of inter-observer variability.

Intra-analyzer reproducibility was defined as the reproducibility calculated by one of the echocardiographer's re-analysis of his own recordings (Z.F.). Inter-analyzer reproducibility was defined as the reproducibility calculated by the two echocardiographers' analyses of the same set of recordings.

Examination of diastolic function. LV mass: LV mass was measured from 2-dimensional (2D) echocardiography, using the recently published guidelines of the American Society of Echocardiography. We measured the septal wall thickness (SWT), posterior wall thickness (PWT) and LV internal diastolic diameter (LVID) from the parasternal 2D long axis view. These measurements were subsequently used in the cube-function formula for ASE guidelines to calculate LV mass [7].

LA Volume: The measurement of left atrial (LA) volume was obtained from the biplane method of disks (modified Simpson's rule) using apical 4-chamber (A4C) and apical 2-chamber (A2C) views at ventricular end systole [7].

Mitral Inflow: The following parameters were studied: peak early filling (E-wave) and late diastolic filling (A-wave) velocities, E/A ratio, deceleration time (DT) of early filling velocity and mitral A-wave duration obtained at the tip of the mitral valve leaflets.

Pulmonary Venous Flow: Measurements included peak systolic (S) velocity, peak antero-grade diastolic (D) velocity, the peak reversal A-wave (Ar) velocity in late diastole and the duration of Ar velocity,

Tissue Doppler Annular Early and Late Diastolic Velocities: Measurements included systolic (S), early diastolic (Ea) and late diastolic velocity (Aa) of the septal and lateral mitral annulus [8].

Assessment of left ventricular diastolic dysfunction was made according to the recommendations of the American Society of Echocardiography and the Committee of the European Association of Echocardiography [3] using the following grading scheme: mild or grade I (impaired relaxation pattern), moderate or grade II (PNF) and severe (restrictive filling) or grade III [3]. Grades II and III were grouped together because of the relatively small number of subjects in these 2 groups.

Assessment of LVEF. LVEF was calculated using the biplane Simpson's method.

Biostatistics methods

The reproducibility of continuous variables is expressed by the intra-class correlation coefficients (ICC). As all analyzed parameters constituted continuous variables, the inter-observer and intra-observer agreements were estimated using the ICC. Agreement was classified as poor ($ICC < 0.40$), fair to good ($ICC = 0.40-0.75$) or excellent ($ICC > 0.75$).

For descriptive purposes, the Bland Altman plots report the means of the ratings for each patient by two raters vs. the differences in ratings. Mean absolute error was calculated as the mean of the absolute difference between the two sets of observations [9]. The mean relative error was calculated as the absolute difference between the two sets of observations divided by the mean of the observations. To assess the degree of agreement between techniques, 95% confidence limits (i.e. mean absolute or relative error $\pm 1.96 \times SD$) were calculated and plotted on each graph.

P values < 0.05 were considered statistically significant. All analyses were performed using SAS version 9.3 (SAS Institute Inc., Cary, N.C., USA) and SPSS 21 (SPSS, Chicago, IL, USA).

Table 1. Patient characteristics.

Characteristics	
Female (%)	33 (55%)
Age (years, mean±SD)	50 ± 14
Height (m, mean±SD)	1.68 ± 0.09
Weight (Kg, mean±SD)	73.55 ± 19.19
BMI (Kg/m ² , mean±SD)	26 ± 6
Current tobacco use (%)	8 (13%)
Hypertension (%)	15 (25%)

Legend: BMI, body mass index.

doi:10.1371/journal.pone.0122336.t001

Results

Overall, our study population was middle-aged, overweight and non-smoking (13%), with a quarter of the participants having self-declared hypertension or being on anti-hypertensive medication (Table 1).

Intra-observer reliability

The intra-observer ICC coefficients were >0.90 for all analyzed parameters except for LVEF which was 0.89 (0.81–0.93) (Table 2).

The mean relative intra-observer differences were <5% for all parameters (Fig 1). The limits of agreement of relative differences for all analyzed parameters were small (2 SDs <15%), indicating a high reproducibility of within-observer measurements. Importantly, the reproducibility of both mitral and Pulmonary A wave duration was excellent (ICC = 0.94 and 0.96 respectively).

Inter-observer reliability

The inter-observer ICC coefficients were >0.90 for most analyzed parameters (Table 2). Slightly lower ICC were observed for LVEF (ICC = 0.87) and both mitral and pulmonary A wave duration (0.83 and 0.73 respectively). Of note, the ICC observed for pulmonary A wave duration was lower than 0.80.

Table 2. Intra- and inter-observer intra-class correlation coefficients for echocardiographic variables.

	Intra-observer ICC		Inter-observer ICC	
	ICC	95% CI	ICC	95% CI
LVEF (%)	0.89	0.81–0.93	0.87	0.78–0.93
LV mass (g)	0.98	0.97–0.99	0.95	0.91–0.97
LA volume (ml)	0.99	0.99–1.00	0.99	0.98–1.00
E wave (m/s)	0.98	0.97–0.99	0.97	0.96–0.99
A wave (m/s)	0.98	0.97–0.99	0.97	0.94–0.98
Septal Ea Velocity (cm/s)	0.99	0.99–1.00	0.98	0.97–0.99
Lateral Ea Velocity (cm/s)	0.99	0.99–1.00	0.99	0.98–1.00
Deceleration time (ms)	0.98	0.97–0.99	0.96	0.94–0.98
A wave duration (ms)	0.94	0.91–0.97	0.83	0.72–0.90
Pulmonary reversal A wave duration(ms)	0.96	0.93–0.98	0.73	0.55–0.84

Legend: ICC, Intra-class correlation coefficients; LVEF, left ventricular ejection fraction; LV, left ventricular; LA, left atrial.

doi:10.1371/journal.pone.0122336.t002




Fig 1. Bland-Altman plots of intra-observer agreement for left ventricular ejection fraction, LV mass, LA Volume, E-wave, E-wave deceleration time, A-wave velocities, A-wave and pulmonary vein reversal A duration, as well as septal and lateral Ep Velocity.

doi:10.1371/journal.pone.0122336.g001

As expected, most limits of agreement were larger for inter-observer comparisons than that observed for intra-observer comparisons (Fig 2). The mean relative inter-observer differences were <5% for all parameters except for the velocity time integral of pulmonary venous flow (mean relative difference = 6.5%). Limits of agreement of relative differences were small (2 SDs <15%), except for mitral A wave duration and velocity (both <20%) as well as for left ventricular mass and pulmonary A wave duration (both <30%).

Intra-observer and inter-observer agreement

Intra-observer agreement with regard to both the presence of diastolic dysfunction and the severity of diastolic dysfunction was excellent (respectively Kappa = 0.93 (0.83–1.00) and 0.88 (0.75–0.99) (Table 3). Inter-observer agreement with regard to the presence of diastolic dysfunction was excellent whereas it was good with regard to the severity of diastolic dysfunction.

Discussion

In the present study, we analyzed the reproducibility of all previously validated Doppler indices of diastolic function. The main finding of this study was that both inter- and intra-observer reproducibility was excellent for most diastolic function parameters with limits of agreement sufficiently narrow to conclude that the fluctuation in measurements of diastolic function is unlikely to have a noteworthy impact on diastolic function classification. We also report excellent inter- and intra-observer agreements with regard to diastolic dysfunction identification and grading. Overall, these results thus qualify the STANISLAS cohort for further research in




Fig 2. Bland-Altman plots of inter-observer agreement for left ventricular ejection fraction, LV mass, LA Volume, E-wave, E-wave deceleration time, A-wave velocities, A-wave and pulmonary vein reversal A duration, as well as septal and lateral Ep Velocity.

doi:10.1371/journal.pone.0122336.g002

the field of diastolic function, including the validation of biomarkers to predict and classify echocardiographic diastolic function.

While prior investigations have assessed the magnitude of variability parameters of diastolic function, these measurements were performed separately and in limited numbers of normal subjects in clinical setting [7,10–13]. The present study is the first to evaluate the reproducibility of all parameters of diastolic function in an unselected population-based cohort leading to a global and better assessment of diastolic function. In addition, the current study is unique in analyzing the variability of Doppler diastolic function indices in a large random sample of subjects in a population-based cohort of initially healthy subjects.

Table 3. Kappa measurements of intra- and inter-observer agreement with regard to diastolic dysfunction.

	Kappa (95% CI)	
	Diastolic dysfunction	Diastolic dysfunction grades (0, 1, ≥2)
Intra-observer	0.93 (0.83–1.00)	0.88 (0.75–0.99)
Inter-observer	0.88 (0.76–1.00)	0.74 (0.59–0.89)

Legend: 95% CI: 95% confidence interval

doi:10.1371/journal.pone.0122336.t003

Intra-observer reliability and agreement

The intra-observer reproducibility for LV systolic function observed in the present study (ICC = 0.89) was globally better than that previously reported. In the study of Palmieri et al., a poor reliability was observed for measurements of LV systolic function (ICC = 0.65) [14]. In contrast, Otterstad et al. demonstrated that differences between repeated echo recordings and repeated video measurements were small and not statistically significant [15]. Furthermore, the findings herein are in agreement with previous studies reporting a high intra-observer reproducibility of diastolic filling and LV mass measurements [7]. In the study of Galderisi et al. [16], intra-observer correlation coefficients were close to 0.90 while the intra-observer percent bias was <10% for Peak velocity of the E and A wave. In contrast with our study, the intra-observer correlation coefficient reported by these authors was <0.90 for deceleration time while the intra-observer relative bias was >10%. Meanwhile, Nagueh et al. [8] reported a mean percentage of intra-observer error of $5\% \pm 4$ for E / e'. In our study, the mean relative intra-observer differences were systematically <5% for these parameters. With regard to LV mass, Grandits et al. [13] reported means and SDs of intra-observer difference in LVM of -0.0 and 20.4 g respectively whereas Palmieri et al. [14] showed an intra-observer difference in LV mass of -1.7 ± 19.8 g for absolute values with a high reliability for LV mass as estimated by an ICC = 0.93. In our study, we detected small means and SDs of absolute intra-observer differences for LV mass (0.4g and 17.5, respectively). Consequently to these results, our measurement protocol enabled a consistent reading of stored echocardiographic data. This very good intra-observer reliability in turn resulted in an excellent agreement with regard to identification and grading of diastolic dysfunction.

Inter-observer reliability and agreement

The present study also yielded superior results than previous studies regarding inter-observer reproducibility of systolic and diastolic function as well as LV mass measurements. Indeed, in the study of Otterstad et al. [15], there was a systematic difference between the two investigators in their measurements of left ventricular ejection fraction. As for diastolic function, Galderisi et al. [16] reported an inter-observer percent bias >10% for time velocity integral E and deceleration time, in addition to a relatively weak inter-observer agreement (ICC = 0.76) for deceleration time. Nagueh et al. [8] reported a mean percentage of inter-observer error of $6\% \pm 5$ for E / e'. In contrast, Kuznetsova et al. [17], upon assessing inter-observer measurements of 17 tracings performed by one experienced recorder, reported an inter-observer mean absolute difference of 0.05 cm/s, 0.02 cm/s, 0.31 ms and 1.42 ms for E wave velocity, A wave velocity, mitral and pulmonary A wave duration, respectively. In our study, the mean relative inter-observer differences were <5% for these parameters, which are in keeping with the results reported by Kuznetsova et al. [17].

LV Linear Dimensions and Wall Thickness are used for virtually all clinical trials that incorporate echocardiography. Linear dimensions are obtainable from correctly-aligned 2D and M-mode images. M-mode recordings provide better temporal resolution for accurate timing of motion of cardiac walls and valves, whereas 2D provides better spatial orientation. With respect to LV mass, Valdez et al. [18] found statistically significant inter-observer differences in measurements of end-diastolic septal thickness, end-diastolic and end-systolic LV posterior wall thicknesses as well as end-diastolic and end-systolic LV diameters by 3 observers on 20 echocardiograms. However, the maximum mean difference was 2 mm which would translate into a moderate error in terms of LV mass. Finally, Grandits et al. [13] reported a mean \pm SD of 7.9 ± 34.7 g for inter-observer difference between the 2 cardiologists. In our study, we detected a small mean of inter-observer difference with moderately wide limits of agreement for

LV mass (<5% and <30% respectively). This overall very good inter-observer reliability resulted in a good to excellent agreement with regard to diastolic dysfunction identification and grading (Kappa = 0.88 (0.76–1.00) and 0.74 (0.59–0.89), respectively). Along with the aforementioned excellent intra-observer agreement, this level of inter-observer agreement qualifies diastolic dysfunction as a relevant and worthy inclusion criterion and outcome in clinical research.

In our study, overall inter-observer agreement in the assessment of diastolic dysfunction classification was better than the results of a multicenter study reported by Unzek et al. [18–21]. In this previous study [18–21], a complete diastolic assessment was obtained from the echocardiographic data of 20 patients and interpreted by 14 experts in 8 countries using the current ASE/EAE recommendations for diastolic assessment [3]. The agreement between raters for assessing diastolic class was moderately lower (Fleiss kappa = 0.62) than that observed in the present study (Kappa = 0.74). In our opinion, the lower agreement reported by Unzek et al. is probably the consequence of the multicenter setting of this previous study, such that the use of the same current ASE/EAE algorithm, even by expert echocardiographers, may still have contained sufficient ambiguity to lead to inconsistent final results. In contrast, in our single-center study, with readers working in the same echocardiography laboratory and applying the same echocardiography working standards, the use of the ASE/EAE algorithm resulted in very similar decision-making processes regarding diastolic dysfunction classification. Yet, if high reproducibility can be achieved in a single center, a harmonization process could be undertaken across centers to achieve a better multi-center agreement with regard to diastolic function grading. Such a process could prove valuable in harmonizing the diagnosis of diastolic dysfunction and heart failure with preserved ejection fraction. However, the modalities of such a harmonization process remains to be determined in further studies.

Large difference in pulmonary A wave duration assessment

The present study revealed a low inter-observer correlation for pulmonary A wave duration with a relatively weak inter-observer agreement (ICC = 0.73) and mean relative inter-observer differences >5%. This result can be explained by the difficulty in the acquisition of this parameter. Indeed, one of the important limitations in interpreting pulmonary venous flow is the difficulty in obtaining high-quality recordings suitable for measurements. This is especially true for pulmonary A wave duration because of low-velocity wall motion artifacts created by atrial contraction that obscure the pulmonary flow velocity signal. Importantly, this variable involved in the determination of filling pressures is not present in the ASE/ESC algorithm of diastolic dysfunction [3]. Consequently, the poor reproducibility of A wave duration would thus impact on the qualification of patients for HF with preserved LVEF rather than on the quantification of diastolic dysfunction according to the algorithm of the ASE/ESC [3].

Differences in imaging validation protocols

Most previously-conducted studies varied considerably in terms of study design and analysis methods, making direct comparison difficult [18–21]. There are many ways to identify as well as control acquisition and measurement variability; however, successful quantification of diastolic function is entirely dependent on image quality. Each echocardiography core lab must be able to demonstrate the efficiency of its work, including operational processes and accuracy of echocardiographic measurement and interpretation. In our study, the physicians and technicians ensured high-quality acquisition data such that accurate quantitative results could be obtained.

Our analysis results compared favorably with other studies. These good results are explained by the design of our protocol based on reanalysis of the same recorded images collected

previously by trained operators with a high-resolution system. According to Otterstad et al. [15], the major component of the random variation in left ventricular volumes and function was repeated echo recordings whereas repeated video measurements and inter-observer variability had a lesser impact. In the present study, we also assessed the magnitude of variability of Doppler diastolic measurements in a relatively large number of healthy patients, which likely increased the reproducibility of our measurements.

Clinical perspectives

The successful application of echocardiography in clinical heart failure trials depends on careful attention to quality control, including echocardiographic acquisition and interpretation harmonization by an experienced and competent core laboratory. The present study thus focused on specific echocardiographic parameters to obtain positive evidence of abnormal LV relaxation, filling, diastolic distensibility and diastolic stiffness, in addition to assessing the reliability of specific parameters needed to classify the patient according to the diagnostic flow-chart of the Echocardiography Associations of the European Society of Cardiology.

Natriuretic peptides and echocardiography are the main tools to detect HF. In the setting of HF with preserved LVEF, echocardiography is of key importance, especially in older patients in whom the use of natriuretic peptides has some limitations. However, access to echocardiography is sometimes difficult and/or long and further compounded by the fact that echocardiographic measurements can show high individual variability in addition to varying day to day in the same patient with changes in preload, afterload and heart rate. We are currently conducting, within the HOMAGE FP7 program (<http://www.homage-hf.eu/>), a biomarker validation (algorithm, program) for the prediction of diastolic dysfunction and presymptomatic heart failure. Obviously, such a validation can solely be obtained for a reproducible echo assessment of diastolic function.

In the current validation study, all of the echocardiographic parameters assessing diastolic function were found to be highly reproducible in the STANISLAS cohort. Given this high reproducibility of diastolic function assessment, and more specifically in this population-based study, we are thus confident of being able to validate a number of new biomarkers to predict diastolic dysfunction.

In addition, the high reproducibility of diastolic function parameters combined with the observed good agreement relative to diastolic dysfunction identification and grading has important practical clinical implications. We provide strong evidence to support that echocardiographic assessment of diastolic function is a reliable and trustworthy clinical tool in terms of metronomic properties.

Limitations

The results of our study are based on patients with normal sinus rhythm, which may limit the generalizability of our findings to subjects in atrial fibrillation. Further studies are warranted to assess the reproducibility of diastolic dysfunction grading in patients with atrial fibrillation.

Conclusion

In the present study, we have validated an operating process for echocardiographic image analysis and quality control measurement for the assessment of diastolic function in patients with sinus rhythm of the STANISLAS cohort study. The excellent results observed in our analysis compared with other studies are explained by the design of our protocol and the single-center nature of our study. This excellent reproducibility will enable the validation of new biomarkers within EU- and French-funded research programs. In addition, in light of these results,

diastolic dysfunction echocardiographic assessment thus appears as a highly reliable clinical and research tool.

Acknowledgments

This study was sponsored by the Nancy CHU. The study was supported by the F-CRIN (F-Clinical Research Infrastructure Network) INI-CRCT (Cardiovascular and Renal Clinical Trialists) network.

The authors deeply thank the entire Clinical Investigation Centre staff all of whom are involved in the daily management of the STANISLAS cohort. We also thank Mr. Pierre Pothier for the editing of the manuscript.

Author Contributions

Conceived and designed the experiments: ZF NG OH FZ PR. Performed the experiments: ZF OH AO EB. Analyzed the data: NG OH ZL. Contributed reagents/materials/analysis tools: ZF OH AO NG PYC. Wrote the paper: ZF NG PR EB AO ZL FZ.

References

1. Zile MR, Baicu CF, Gaasch WH. Diastolic heart failure—abnormalities in active relaxation and passive stiffness of the left ventricle. *N Engl J Med*. 2004; 350: 1953–1959. PMID: [15128895](#)
2. Adams K, Lindenfeld J, Arnold JMO, Baker DW, Barnard DH, Baughman KL et al. Executive summary: HFSA 2006 Comprehensive Heart Failure Practice Guideline. *J Card Fail*. 2006; 12: 10–38. PMID: [16500578](#)
3. Nagueh SF, Appleton CP, Gillebert TC, Marino PN, Oh JK, Smiseth OA, et al. Recommendations for the evaluation of left ventricular diastolic function by echocardiography. *Eur J Echocardiogr*. 2009; 10: 165–193. doi: [10.1093/ejechocard/jep007](#) PMID: [19270053](#)
4. Hatle L. How to diagnose diastolic heart failure a consensus statement. *Eur Heart J*. 2007; 28: 2421–2423. PMID: [17890729](#)
5. Paulus WJ, Tschope C, Sanderson JE, Rusconi C, Flachskampf FA, Rademakers FE, et al. How to diagnose diastolic heart failure: a consensus statement on the diagnosis of heart failure with normal left ventricular ejection fraction by the Heart Failure and Echocardiography Associations of the European Society of Cardiology. *Eur Heart J*. 2007; 28: 2539–2550. PMID: [17428822](#)
6. Last JM. *A Dictionary of Epidemiology*. New York: NOUP.
7. Lang RM, Bierig M, Devereux RB, Flachskampf FA, Foster E, Pellikka PA, et al. Recommendations for chamber quantification. *Eur J Echocardiogr*. 2006; 7: 79–108. PMID: [16458610](#)
8. Nagueh SF, Middleton KJ, Kopelen HA, Zoghbi WA, Quinones MA. Doppler tissue imaging: a noninvasive technique for evaluation of left ventricular relaxation and estimation of filling pressures. *J Am Coll Cardiol*. 1997; 30: 1527–1533. PMID: [9362412](#)
9. Bland JM, Altman DG. Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet*. 1986; 1: 307–310. PMID: [2868172](#)
10. Spirito P, Maron BJ, Verter I, Merrill JS. Reproducibility of Doppler echocardiographic measurements of left ventricular diastolic function. *Eur Heart J*. 1988; 9: 879–886. PMID: [3181174](#)
11. Valantine HA, Hatle LK, Appleton CP, Gibbons R, Popp RL. Variability of Doppler echocardiographic indexes of left ventricular filling in transplant recipients and in normal subjects. *J Am Soc Echocardiogr*. 1990; 3: 276–284. PMID: [2206544](#)
12. Kuecherer HF, Kee LL, Modin G, Cheitlin MD, Schiller NB. Echocardiography in serial evaluation of left ventricular systolic and diastolic function: importance of image acquisition, quantitation, and physiologic variability in clinical and investigational applications. *J Am Soc Echocardiogr*. 1991; 4: 203–214. PMID: [1854491](#)
13. Grandits GA, Liebson PR, Dianzumba S, Prineas RJ. Echocardiography in multicenter clinical trials: experience from the Treatment of Mild Hypertension Study. *Control Clin Trials*. 1994; 15: 395–410. PMID: [8001359](#)
14. Palmieri V, Dahlof B, DeQuattro V, Sharpe N, Bella JN, de Simone G, et al. Reliability of echocardiographic assessment of left ventricular structure and function: the PRESERVE study. Prospective

- Randomized Study Evaluating Regression of Ventricular Enlargement. *J Am Coll Cardiol.* 1999; 34: 1625–1632. PMID: [10551715](#)
15. Otterstad JE, Froeland G, St John Sutton M, Holme I. Accuracy and reproducibility of biplane two-dimensional echocardiographic measurements of left ventricular dimensions and function. *Eur Heart J.* 1997; 18: 507–513. PMID: [9076390](#)
 16. Galderisi M, Benjamin EJ, Evans JC, D'Agostino RB, Fuller DL, Lehman B, et al. Intra- and interobserver reproducibility of Doppler-assessed indexes of left ventricular diastolic function in a population-based study (the Framingham Heart Study). *Am J Cardiol.* 1992; 70: 1341–1346. PMID: [1442588](#)
 17. Kuznetsova T, Herbots L, Lopez B, Jin Y, Richart T, Thijs L, et al. Prevalence of left ventricular diastolic dysfunction in a general population. *Circ Heart Fail.* 2009; 2: 105–112. doi: [10.1161/CIRCHEARTFAILURE.108.822627](#) PMID: [19808325](#)
 18. Valdez RS, Motta JA, London E, Martin RP, Haskell WL, Farquhar JW, et al. Evaluation of the echocardiogram as an epidemiologic tool in an asymptomatic population. *Circulation.* 1979; 60: 921–929. PMID: [476895](#)
 19. Unzek S, Popovic ZB, Marwick TH, Diastolic Guidelines Concordance I. Effect of recommendations on interobserver consistency of diastolic function evaluation. *JACC Cardiovasc Imaging.* 2011; 4: 460–467. doi: [10.1016/j.jcmg.2011.01.016](#) PMID: [21565732](#)
 20. Ladipo GO, Dunn FG, Pringle TH, Bastian B, Lawrie TD. Serial measurements of left ventricular dimensions by echocardiography. Assessment of week-to-week, inter- and intraobserver variability in normal subjects and patients with valvular heart disease. *Br Heart J.* 1980; 44: 284–289. PMID: [7426186](#)
 21. Schieken RM, Clarke WR, Mahoney LT, Lauer RM. Measurement criteria for group echocardiographic studies. *Am J Epidemiol.* 1979; 110: 504–514. PMID: [507041](#)