



HAL
open science

Bone drilling: a thermal model for bone temperature prediction

Foli Amewoui, Gael Le Coz, Anne Sophie Bonnet, Abdelhadi Moufki

► **To cite this version:**

Foli Amewoui, Gael Le Coz, Anne Sophie Bonnet, Abdelhadi Moufki. Bone drilling: a thermal model for bone temperature prediction. *Computer Methods in Biomechanics and Biomedical Engineering*, 2019, 22, pp.S305-S307. 10.1080/10255842.2020.1714922 . hal-03040617

HAL Id: hal-03040617

<https://hal.univ-lorraine.fr/hal-03040617>

Submitted on 4 Dec 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Bone drilling: a thermal model for bone temperature prediction

F. Amewoui, G. Le Coz, A. S. Bonnet and A. Moufki

Laboratoire d'Etude des Microstructures et de Mécanique des Matériaux, LEM3, Université de Lorraine, CNRS, Arts et Métiers Paris Tech, Metz, France

1. Introduction

Bone drilling is a common operation in orthopaedic and dental practices. This operation generates heat resulting mainly from friction between bone hole wall, chips and drill bit body. Heat generation during bone drilling exposes bone cells to elevated temperature and may result in thermonecrosis. Although many experimental investigations have been conducted to characterize the optimal drilling conditions, conclusions are still uncertain and diverging (Pandey and Panda 2013). In the present study, a semi-analytical model, based on the moving heat source approach combined with the method of image sources, is developed for predicting the temperature during bone drilling. An experimental drilling campaign on porcine bone samples is conducted in parallel, to validate the model.

2. Methods

2.1. Bone temperature rise prediction model

To model the bone temperature rise, two heat sources are considered, see Figure 1. First the interaction between the drill-bit tip and the bone is modelled as a disk heat source. The second source, due to the friction between the drill-bit side and the hole wall, is assimilated to a finite number of moving ring sources which are activated progressively as the drill-bit body penetrates the bone. The disk heat source is assumed to be equal to a fraction $0 < \eta < 1$ of the mean value of the mechanical energy:

$$q_c = \eta(\bar{F}V + \bar{M}\omega)/(\pi R^2) \quad (1)$$

where \bar{F} and \bar{M} represent respectively the average of the maximum and minimum values of feed force and cutting torque measured experimentally. V is the feed

rate, ω is the rotational speed and R is the tool radius.

The frictional heat induced by the contact between the drill-bit side and the drilled hole is given by:

$$q_f = \mu_f p V_c \quad (2)$$

with μ_f the friction coefficient ($=0.3$), p the contact pressure and V_c the cutting velocity ($V_c = \omega R$).

Bone sample is assimilated to a half space. To get the temperature distribution in the semi-infinite medium with adiabatic surface, image source approach is introduced. By using the works of Carslaw and Jaeger (1959), the temperature elevation, at time t , at point $M(x, y, z)$, due to the amount of heat qds liberated at time t' at point $Q(x', y', z')$, by an elementary heat source moving along x -direction with constant velocity V , is given by:

$$\Delta T(x, y, z, t) = \frac{qds}{8\rho c(\pi a)^{3/2}} \int_0^t (t-t')^{-3/2} e^{-d^2/4a(t-t')} dt'$$

$$with d^2 = (x - Vt')^2 + (y - y')^2 + (z - z')^2 \quad (3)$$

The values of thermal conductivity, specific heat and apparent density are respectively (Tu et al. 2012): $k = 0.45 \text{ W/mK}$, $c = 1640 \text{ J/kgK}$, $\rho = 1640 \text{ kg/m}^3$. The thermal diffusivity is denoted by a .

The temperature distribution due to the drill-bit tip contribution is obtained by adding the temperature rise due to each elementary heat source over the disk surface. The effect of friction at the drill-bit side is obtained by integrating the elementary heat source over each moving ring progressively activated. Finally, the resulting temperature distribution is given by:

$$T(x, y, z, t) = T_0 + T_{tip}(x, y, z, t) + T_{side}(x, y, z, t) \quad (4)$$

with T_0 the initial bone temperature. The values of model parameters $\eta (=0.35)$ and $p (1 \text{ MPa})$ were calibrated with respect to the smallest cutting speed $V_c = 2 \text{ m/min}$.

2.2. Experimental study

The bone samples were obtained by slicing, into lamella of about $70 \text{ mm} \times 10 \text{ mm}$, the diaphysis of a porcine femur provided by a local slaughter. The bone specimens were frozen at -20°C and they were




Figure 1. Illustration of the two heat sources considered in the model.




Figure 2. (a) Predicted and measured temperature for the thermocouple located at 0.5 mm from hole wall (b) Comparison of maximal temperature elevation.

thawed at ambient temperature 24 hours prior to the tests. Drilling experiments were conducted on a three axes CNC milling machine Roeders RP600. Nobel Biocare REF32271 twist drills with a diameter of 3.2 mm, a helix angle of 30° and point angle of 118° were used. Two K-type thermocouples OMEGA

KMTSS-IM025U-150 with a wire diameter of 0.25 mm allowed to measure the temperature elevation inside bone. They were inserted into two holes located at 0.5 mm and 0.9 mm from the hole wall. No lubrication was used. The feed force and drilling moment were also recorded to estimate the heat source at the drill tip (see Equation (1)). The measurements were done thanks to a Kistler 9256C2 piezoelectric dynamometer. The drilling experiments were performed by fixing the feed per tooth at 0.02 mm and varying cutting velocities between 2 and 20 m/min at room temperature. The drilling depth was 4 mm.

3. Results and discussion

After model calibration, the temperature predicted by the model and obtained by measurement are compared (Figure 2(a)). Two phases are visible on these curves: the first one corresponding to the increase of temperature during the drilling process and the second one showing the cooling stage. Although some improvements need to be made in order to better describe the cooling phase, the profiles of temperature elevations are globally reproduced by the model. Concerning the comparison of the maximal temperature rises illustrated by Figure 2(b), the increase of temperature with cutting velocity from 2 to 10 m/min is properly described. From 10 to 20 m/min, a diminution is observed experimentally while the temperature predicted by the model tends to saturate. These temperature elevations reach values included between 25 and 60 degrees, but no conclusion can be drawn concerning a possible necrosis of bone cells, as no lubrication was taken into account in this work.

A constant value of bone apparent density was considered in this model but XRay tomography can be carried out in order to determine the local value of density and adjust the values of thermal properties, according to the variation of bone density. Indeed, the effect of this quantity on heat generation is significant and has to be considered, (Feldmann et al. 2017).

4. Conclusions

A semi-analytical model based on the moving heat source approach has been developed to reproduce the bone temperature variation during drilling. Improvements must be made to take into account the influence of bone microstructure on heat generation but the model already provides a good estimation of the temperature rise.

References

- Carslaw HS, Jaeger JC. 1959. Conduction of heat in solids. Oxford University Press.
- Feldmann A, Gavaghan K, Stebinger M, Williamson T, Weber S, Zysset P. 2017. Real time prediction of temperature elevation during robotic bone drilling using the torque signal. *Ann Biomed Eng.* 45(9):2088–2097.
- Pandey RK, Panda SS. 2013. Drilling of bone: a comprehensive review. *J Clin Orthop Trauma.* 4(1):15–30.
- Tu YK, Chen LW, Ciou JS, Hsiao CK, Chen YC. 2012. Finite element simulations of temperature rise during bone drilling based on a bone analog. *J Med Biol Eng.* 33:269–274.

KEYWORDS Bone drilling; thermonecrosis; thermal model; moving heat source

 anne-sophie.bonnet@univ-lorraine.fr